



Correction to: Simultaneous detection of severe acute respiratory syndrome, Middle East respiratory syndrome, and related bat coronaviruses by real-time reverse transcription PCR

Ji Yeong Noh^{1,2} · Sun-Woo Yoon^{1,3} · Doo-Jin Kim¹ · Moo-Seung Lee^{1,3} · Ji-Hyung Kim¹ · Woonsung Na⁴ · Daesub Song⁴ · Dae Gwin Jeong^{1,3} · Hye Kwon Kim¹

Published online: 22 December 2017
© Springer-Verlag GmbH Austria, part of Springer Nature 2017

Correction to: Arch Virol (2017) 162:1617–1623
<https://doi.org/10.1007/s00705-017-3281-9>

Unfortunately, the concentration unit of plasmids was published incorrectly in the original publication of the article. The concentration unit, ‘copies/ml’ should be corrected to ‘copies/ μ l’. This changes do not affect to the analytic sensitivity of the method because the detection limits of 50–100 copies/ μ L and 5–100 copies/ μ L using pUC57-SARS-pS2 (a template for SARS-CoV) and

pGEM-MERS-S2 (a template for MERS-CoV), respectively, were as sensitive as other real-time PCR methods [1].

Reference

1. Parker J, Fowler N, Walmsley ML, Schmidt T, Scharrer J, Kowaleski J, Grimes T, Hoyos S, Chen J (2015) Analytical sensitivity comparison between singleplex real-time PCR and a multiplex PCR platform for detecting respiratory viruses. *PLoS One* 10:e0143164

The original article can be found online at <https://doi.org/10.1007/s00705-017-3281-9>.

✉ Dae Gwin Jeong
dgjeong@kribb.re.kr

✉ Hye Kwon Kim
khk1329@kribb.re.kr

- 1 Infectious Disease Research Center, Korea Research Institute of Bioscience and Biotechnology, Daejeon, Republic of Korea
- 2 College of Veterinary Medicine, Chungbuk National University, Cheongju 28644, Republic of Korea
- 3 Bio-Analytical Science Division, Korea University of Science and Technology (UST), Daejeon, Republic of Korea
- 4 Department of Pharmacy, College of Pharmacy, Korea University, Sejong, Republic of Korea