



# An integrated target field framework for point-of-care halbach array low-field MRI system design

Bart de Vos<sup>1</sup> · Rob F. Remis<sup>2</sup> · Andrew G. Webb<sup>1,2</sup>

Received: 14 November 2022 / Revised: 18 January 2023 / Accepted: 16 April 2023 / Published online: 20 May 2023  
© The Author(s) 2023

## Abstract

**Objective** Low-cost low-field point-of-care MRI systems are used in many different applications. System design has correspondingly different requirements in terms of imaging field-of-view, spatial resolution and magnetic field strength. In this work an iterative framework has been created to design a cylindrical Halbach-based magnet along with integrated gradient and RF coils that most efficiently fulfil a set of user-specified imaging requirements.

**Methods** For efficient integration, target field methods are used for each of the main hardware components. These have not been used previously in magnet design, and a new mathematical model was derived accordingly. These methods result in a framework which can design an entire low-field MRI system within minutes using standard computing hardware.

**Results** Two distinct point-of-care systems are designed using the described framework, one for neuroimaging and the other for extremity imaging. Input parameters are taken from literature and the resulting systems are discussed in detail.

**Discussion** The framework allows the designer to optimize the different hardware components with respect to the desired imaging parameters taking into account the interdependencies between these components and thus give insight into the influence of the design choices.

**Keywords** Low-Field MRI · System Design · Inverse source problem · Halbach array · Gradient coil · RF coil

## Introduction

Access to MRI is severely limited in many countries [1, 2]. According to the International Atomic Energy Agency the number of MRI's per million inhabitants in Sub-Saharan Africa is 0.3, while this figure is 22.2 averaged throughout Europe and Northern America [3]. Developing low-field strength ( $\ll 1$  Tesla) systems has the potential of improving this situation, decreasing the overall cost by eliminating the need for an expensive superconducting magnet, high-power water-cooled gradient coils, proprietary and hard-to-fix electronics, and elaborate site preparation. Accessibility can also be improved by designing these low-field systems such that

they can be brought to the patient (point-of-care systems). Several recent reviews on the topic of low-field MRI have been published [4–8].

Multiple groups have successfully performed *in-vivo* experiments with point-of-care systems. In the summary below, individual image datasets were typically acquired in  $\sim 10$  min. Popular point-of-care magnet designs are the H- and C-type parallel plate systems, where the letter refers to the shape of a yoke made of a material with high permeability, connecting the plates containing the permanent magnets. Nakagomi et al. [9] designed and constructed an H-type magnet weighing 200 kg, with an average field strength of 200 mT measured in a 100 mm diameter spherical volume (DSV). The system was mounted and operated from inside a mini-van. Elbow images were obtained using a field-of-view (FOV) of  $180 \times 180$  mm<sup>2</sup> and a digital resolution of  $\sim 0.7 \times 1$  mm<sup>2</sup>, using 9 slices with a thickness of 3 mm. He et al. [10] created a similar H-type system for a larger imaging volume (200 mm DSV), and showed *in-vivo* images of stroke patients using a 350 kg, 50.9 mT magnet. The images obtained with this system had a slice thickness of 10 mm and the FOV used was  $260 \times 260$  mm<sup>2</sup> with a resolution of  $\sim 1.5 \times 2$  mm<sup>2</sup>.

✉ Bart de Vos  
b.de\_vos@lumc.nl

Andrew G. Webb  
a.webb@lumc.nl

<sup>1</sup> C.J. Gorter MRI Center, Leiden University Medical Center, Leiden, Netherlands

<sup>2</sup> Signal Processing Systems, Delft University of Technology, Delft, Netherlands

The company Hyperfine has built an FDA approved 64 mT, C-type system, with the magnet weighing 320 kg. The reported obtainable resolution is  $\sim 1.5 \times 1.5 \times 5 \text{ mm}^3$ . The system has been used at the bedside for adult neuroimaging [11–13] and paediatric neuroimaging [14], it has been operated from within a mini-van [15], and has been used in a resource-constrained environment [16]. Using an open-access dataset the diagnostic performance of this device has been simulated and compared to a clinical 3 T system [17]. A similar system for adult neuroimaging was designed recently by Liu et al. [18], which acquired all images with a digital resolution of approximately  $2 \times 2 \times 10 \text{ mm}^3$ . The system without amplifiers weighed 750 kg, and had a field strength of 55 mT measured over a 240 mm DSV. Both the Hyperfine system and the system designed by Liu et al. implemented multiple external antennas which were used to reduce the effects of any electromagnetic interference. For educational purposes, rather than imaging a body region Cooley et al. [19] designed a table top parallel plate system with a 1 cm FOV. Reproducibility and value beyond medical applications were demonstrated by building 20 of these systems.

Other groups have proposed more lightweight solutions using a magnet based on a discrete Halbach array. These systems consist of rings holding thousands of small magnets positioned in a configuration first described by Halbach [20–22], and creating a  $B_0$  field transverse to the bore. Blümner [23–26] created multiple magnet systems using an array of identical bar magnets oriented such that it resembles the magnetisation described by Halbach. Cooley et al. [27, 28] designed an MRI system weighing 280 kg with a gradient in the main magnetic field, and an average field strength of 80 mT measured over a 200 mm DSV. Brain images with a resolution of  $2.2 \times 1.3 \times 6.8 \text{ mm}^3$  were acquired with a FOV of  $200 \times 200 \times 200 \text{ mm}^3$ . Electromagnetic interference was reduced using multiple noise antenna's and an algorithm that uses the impulse response of the sensor data to remove the interference in k-space [29]. In a more recent study the visibility of lesions in the brain were compared to a 1.5 T clinical system [30]. O'Reilly et al. [31, 32] designed a magnet weighing approximately 75 kg with a field strength of 50 mT measured over a 200 mm DSV. Conventional Fourier encoding methods using three gradients were employed. Images of the knee with 2 mm isotropic resolution and brain images with a resolution of  $4 \times 4 \times 4 \text{ mm}^3$  were shown. A similar 70 mT, 200 kg magnet system designed for extremity imaging was created by Guallart-Naval et al. [33]. Wrist and hand images were acquired using a  $200 \times 160 \times 80 \text{ mm}^3$  FOV with a digital resolution of  $0.8 \times 1.3 \times 5 \text{ mm}^3$ .

There are relative advantages and disadvantages associated with each of the approaches to magnet design. A parallel plate design is less involved since it consists simply of large pieces of permanent magnet material and a yoke. In addition, it is more open than the Halbach array which

is a closed cylinder. However, the Halbach system produces the maximum field strength per unit weight of magnetic material and construction is considered to be safer due to the magnetic forces being divided over thousands of small magnets. Gradient coils for the parallel plates must be planar in nature, which are intrinsically less efficient than the cylindrical coils used in a Halbach array. Radiofrequency (RF) coil design is similar, although coupling to the gradient coils is higher in the Halbach design due to the cylindrical geometry.

The overall system dimensions are crucial to facilitate point-of-care since designing the magnet, gradient and RF coils to be as compact as possible is essential for portability. The minimum physical dimensions are ultimately governed by the targeted application. However, there are well known trade-offs between size and performance. Magnets with a smaller diameter bore produce higher magnetic field strengths and are lighter, but have a poorer  $B_0$ -homogeneity over a given field-of-view [31]. Gradient coils with smaller diameter have higher efficiency, and a smaller length/diameter ratio results in poorer linearity [34]. Higher image spatial resolution requires higher gradient strengths with respect to the  $B_0$ -inhomogeneity. Similarly, RF coils with smaller length/diameter have higher transmit/receive efficiency, but poorer RF homogeneity [35].

In this work we introduce an iterative design approach which integrates the different hardware designs (magnet/gradient coils/RF coil) into one framework. Using time efficient methods allows the designer to obtain a fast understanding of the interdependencies of the input parameters to change to obtain desired imaging specifications. As a consequence the user can design optimal point-of-care systems, where optimal corresponds to the most compact solution, which is able to obtain the desired image characteristics with the maximum  $B_0$  field strength for the highest SNR. Doing this via a linear design approach, where the components are designed individually, is sub-optimal due to the many interdependencies. Therefore, we propose a framework which iterates target field methods for gradient, RF and magnet design, where desired fields are prescribed in a region of interest and the inverse source problem is solved in a semi-analytical manner. The advantage of target-field methods is that they have short computation times and can be programmed on a single platform: such methods have not been shown for magnet design, but are often used in gradient, RF- and shim coil design [36–41]. The framework incorporates the interdependencies of the hardware components and their effects on the imaging parameters and uses the gradient and RF amplifier characteristics to create a system which is as compact as possible while corresponding to pre-set imaging requirements. Halbach array based magnets are considered, where ring-diameter and ring-spacing are optimised for homogeneity and where the individual magnet

size, remanence and spacing between them are fixed quantities defined by the user.

### Methods

First the framework is discussed followed by a detailed description of the individual design methods of which it consists.

### Target field framework for iterative system design

A flowchart of the complete design framework is shown in Fig. 1. The computational blocks are connected with arrows where the green arrows denote the natural flow of the framework. The orange arrows show the iterative paths which are taken to revisit certain design blocks to improve the overall design. The red arrows shows when a certain limit defined by the user is reached. This will terminate the design loop. Finally the black arrows show the flow of parameters from one design block to the next. These parameters are specified in Table 1.

Starting at the top left of the framework, the user specifies the input parameters for RF, magnet and gradient design, together with the desired FOV and spatial resolution. Given these, a system is designed which is as compact as possible, has the strongest possible magnetic field and has an effective resolution corresponding to the imaging requirements.


The first computation is that of the initial magnet design. The magnet systems considered are variable ring-diameter

and ring-spacing Halbach arrays, where the individual magnets are placed in discrete rings. Such a setup is shown in Fig. 2. The initial number of rings is set by the user and is fixed within the magnet optimisation algorithm. The other starting conditions are the minimum bore diameter, minimum target field strength, volume of interest and the remanence and spacings of the individual magnets: these are all fixed throughout the framework. Based on these parameters the diameters and spacings of the rings are optimised to approach the user-specified target field while maximising homogeneity.

Next, power optimised gradient coils are designed. Their lengths and radii are determined by the magnet dimensions from the previous magnet optimisation. The most efficient coils are determined given a user-specified linearity.

Subsequently, the gradient-limited effective spatial resolution in the readout direction  $\Delta z_{eff}$  is determined using the efficiency of the least efficient gradient coil, the maximum gradient amplifier current, the  $B_0$  field inhomogeneity of the magnet, and the field-of-view and number of complex data points defined by the user.

If the effective resolution is larger than the predefined threshold the homogeneity of the magnet needs to be increased. In this case the model loops back to the magnet design, and an additional ring is added. The ring-diameters and separations are again optimised creating a more homogeneous (but larger) magnet. The gradient coils are then redesigned given the new magnet geometry. Adding rings increases the allowable length of the gradient coils which increases the linearity. However, the efficiency given




**Fig. 1** Flow chart of the system design framework. The black rectangles represent the input parameters. The light blue blocks hold the different computational methods, the dark blue rectangles positioned inside the methods show the output parameters, the diamonds

represent the if-statements and the circles express the actions exerted by the framework. The green arrows indicate the natural flow of the framework, the orange arrows indicate feedback. Red arrows indicate when interaction with the user is required

**Table 1** Input parameters framework

Quantity	Head system	Extremity system
<b>Imaging parameters</b>		
Target resolution, $\Delta z_T$ [mm]	3×3×3	1.75×1.75×1.75
Target resolution margin, $\epsilon_m$ [mm]	0.1	0.01
Digital resolution [mm]	2.5×3×3	1.60×1.75×1.75
Field-of-view [mm <sup>3</sup> ]	200×200×200	200×120×120
Data matrix	80×66×66	125×68×68
<b>Magnet parameters</b>		
Target Volume [mm <sup>3</sup> ]	200 (DSV)	60×200 (r×L Cylinder)
Minimum magnet radius, $r_{\min}$ [mm]	150	100
Maximum magnet length, $L_{\max}$ [mm]	550	500
Minimum field strength [mT]	45	80
Number of radial magnet layers, $L$	2	2
Magnet size [mm <sup>3</sup> ]	12	12
Azimuthal magnet spacing, $\kappa$ [mm]	20	20
Radial magnet spacing, $\delta$ [mm]	20	20
Magnet remanence, $M_0$ [T]	1.3	1.3
<b>Gradient parameters</b>		
Target Volume [mm <sup>3</sup> ]	200 (DSV)	60×200 (r×L Cylinder)
Max grad amplifier current [A]	30	30
Gradient wire diameter [mm]	1.5	1.5
Min wire spacing [mm]	1	1
Max linearity error, $\sigma$	5%	5%
<b>RF parameters</b>		
Target Volume [mm <sup>3</sup> ]	200 (DSV)	60×200 (r×L Cylinder)
RF coil length [mm]	240	240
RF coil diameter [mm]	240	140
Max RF uniformity error	20%	20%
Min RF pulse length [ $\mu$ s]	50	50
RF amplifier gain [dB]	54	54
Max spectrometer output [dBm]	0	0

**Fig. 2** Illustration of a Halbach array depicting the coordinate system and symbols used. Left: front view of a single ring with two magnet layers. Right: side view of two rings

a certain linearity has an optimum. If increasing the length of the coils during the framework iterations decreases the efficiency, the previous shorter solution is taken.

The process of magnet and gradient design continues to iterate until the requirements for spatial resolution are met. At this point, the framework determines if it is possible to increase the target magnetic field strength  $B_T$ , while maintaining the required spatial resolution. Increasing  $B_T$  effectively decreases the average ring-diameter and spacing, resulting in a smaller but less homogenous magnet.

This iterative process of changing the magnet design by adding rings or increasing the field strength, computing the resulting gradient coils and determining the effective resolution, is repeated until the desired effective spatial resolution is reached with the strongest possible field strength.

In the final step a solenoidal RF coil with variable wire spacing is designed: a solenoid is chosen due to the intrinsically higher sensitivity compared to saddle coils. The coil dimensions are defined by the user and optimised for transmit efficiency given a user-specified non-uniformity over the FOV. The checks performed for the RF coil are twofold: the user defines a minimum pulse length (microseconds) and

the framework ensures that this excites all of the hydrogen nuclei (with a range of resonance frequencies dictated by the  $B_0$  field inhomogeneity) within the imaging FOV. In addition, the power available from the RF amplifier combined with the transmit efficiency of the coil, given the user-specified minimum pulse duration, must be sufficient to obtain the desired flip angles. If these conditions are met the framework is finished.

The framework terminates prematurely if: (i) the pre-specified maximum length of the system ( $L_{magnet} > L_{max}$ ) is exceeded without achieving the requested effective resolution, (ii) if the required RF power is not sufficient to obtain the desired flip angle, or (iii) if the RF pulse is not able to excite the entire FOV. The user can utilize this information to alter the input parameters such that the design is feasible.

### Target field approach for halbach array based magnets with variable ring-diameters and -spacing

In this section, we derive an expression to describe the field created by magnets in an Halbach array with variable ring-diameter and ring-spacing. Determining the diameters and spacing such that a given prescribed target field is realized is a non-linear inverse source problem which is solved using a Newton optimisation scheme.

The individual magnets of the Halbach array are modelled as magnetic dipoles. Many authors have shown that such an approximation provides a sufficiently accurate magnetic flux density approximation within the domain enclosed by the magnet array [27, 31, 42]. The magnetic flux density  $\mathbf{B}$  at position  $\mathbf{x}$  due to a dipole positioned at location  $\mathbf{x}'$  is given by

$$\mathbf{B}(\mathbf{x}) = \frac{\mu_0}{4\pi} \left[ \frac{3(\mathbf{x} - \mathbf{x}') [\mathbf{M} \cdot (\mathbf{x} - \mathbf{x}')]}{|\mathbf{x} - \mathbf{x}'|^5} - \frac{\mathbf{M}}{|\mathbf{x} - \mathbf{x}'|^3} \right], \quad (1)$$

where  $\mathbf{M} = \mathbf{M}(\mathbf{x}')$  is the magnetization vector. In Fig. 2, the magnet configuration is shown including the coordinate system and the various symbols that are used for the design of the magnet: a right handed system is used with  $z$  directed parallel to  $B_0$  and  $x$  parallel to the axis of the bore.

As the optimisation is defined with respect to the ring-diameters and -spacing, cylindrical coordinates  $\mathbf{x}' = \{r', \varphi', x'\}$  are used to indicate the position of the magnets. Furthermore, the magnets are oriented such that a set of magnets within a ring form a discrete approximation of the continuous Halbach magnetization

$$\mathbf{M} = M_0 [\cos(2\varphi') \hat{\mathbf{z}} + \sin(2\varphi') \hat{\mathbf{y}}], \quad (2)$$

where  $M_0$  is the remanent magnetisation of the individual magnets. This magnetization profile creates a homogenous field in the  $z$ -direction in the case of an infinitely long cylinder [20–22]. The magnetization vector of Eq. (2) is now

substituted into Eq. (1), and only the relevant  $z$ -component of the magnetic flux density  $\mathbf{B}$  is considered. Assuming that interactions between the magnets (i.e. any demagnetization effects) can be neglected, the field generated by the array can be found by summing the fields of each individual magnet.

The magnets are located in  $N$  discrete rings and each ring may consist of  $L$  concentric layers of magnets to accommodate for higher field strengths. The radius of the  $l$ th layer in the  $n$ th ring is denoted by  $r_{ln} = r_n + \delta(l - 1)$ , where  $\delta$  is the radial separation between the layers. The azimuthal distance between the magnets inside a layer  $\kappa$  is held constant, meaning that the number of magnets per layer can be found as  $M_{ln} = \frac{2\pi r_{ln}}{\kappa}$ , and therefore varies as a function of the ring-radius. The azimuthal angle of the  $m$ th magnet is  $\varphi_m = m \frac{2\pi}{M_{ln}}$  and we find that.

$$B_z(\mathbf{x}) = \frac{\mu_0}{4\pi} \sum_{n=1}^N \sum_{l=1}^L \sum_{m=1}^{M_{ln}} \left\{ \frac{3(z - z'_{lmm}) [M_z(z - z'_{lmm}) + M_y(y - y'_{lmm})]}{|\mathbf{x} - \mathbf{x}'|^5} - \frac{M_z}{|\mathbf{x} - \mathbf{x}'|^3} \right\}, \quad (3)$$

where  $z'_{lmm} = r_{ln} \cos \varphi'_m$  and  $y'_{lmm} = r_{ln} \sin \varphi'_m$ . Using Eq. (3), the ring-diameters and -spacing can now be optimised using a target field approach. The procedure for the ring-radii is illustrated here. Ring-spacing optimisation is done in a similar manner and is therefore only briefly discussed. The desired target field values at the locations  $[\mathbf{x}_1, \mathbf{x}_2, \dots, \mathbf{x}_K]$  are stored in the  $K \times 1$  vector  $\mathbf{b}_T$ .

The unknown ring-radii are represented by the  $N \times 1$  vector  $\mathbf{r}$  and the  $N \times 1$  vector  $\mathbf{d}$  contains the ring-positions. Finally,  $\mathbf{b}_z(\mathbf{r})$  is the  $K \times 1$  vector obtained by evaluating (3) for every target field position and parameter vector  $\mathbf{r}$ . The objective now is to find vector  $\mathbf{r}$  such that

$$\mathbf{b}_T - \mathbf{b}_z(\mathbf{r}) = 0. \quad (4)$$

This is clearly a non-linear problem, and a Newton optimization scheme [43] is chosen to find a vector  $\mathbf{r}$  for which Eq. (4) is (approximately) satisfied. Specifically, starting with an initial guess  $\mathbf{r}_0$ , the optimal vector is found in an iterative manner using the update equation

$$\mathbf{r}_{i+1} = \mathbf{r}_i + \mathcal{J}_r^+ [\mathbf{b}_T - \mathbf{b}_z(\mathbf{r}_i)]. \quad (5)$$

Here  $\mathcal{J}_r^+$  is the pseudo inverse of the  $K \times N$  Jacobian matrix which is given by

$$\mathcal{J}_r = \begin{bmatrix} \frac{\partial B_z(\mathbf{x}_1, r_1)}{\partial r_1} & \dots & \frac{\partial B_z(\mathbf{x}_1, r_N)}{\partial r_N} \\ \vdots & \ddots & \vdots \\ \frac{\partial B_z(\mathbf{x}_K, r_1)}{\partial r_1} & \dots & \frac{\partial B_z(\mathbf{x}_K, r_N)}{\partial r_N} \end{bmatrix}, \quad (6)$$

To obtain practical solutions for the clear bore of the magnet, lower bounds on the ring-radii are imposed and the



radii are set to the predefined minimum value if the condition is violated. This minimum value is also taken as an initial guess for all ring -radii and defines the starting vector  $\mathbf{r}_0$ . In the case of a homogenous target field positioned at the centre of the magnet, mirror symmetry of the ring diameters and distances can be utilized with respect to the  $x = 0$  plane. Finally, the iterative process is terminated if the relative error falls below a specified tolerance or a maximum number of iterations is reached.

The ring-distance is optimised in a similar fashion, only requiring Eqs. (4–6) to be redefined in terms of the ring-positions  $x'$ . Lower bounds on the ring distance are applied to take the thickness of the rings into account. To fix the magnets in the rings plastic lids are typically used: the minimum practical thickness is  $\sim 0.5$  mm for each lid. Therefore a spacing of 1 mm between the rings is chosen as a starting point.

Because the field strength is more sensitive to small changes in the ring-radii than to changes in the ring-spacing [25, 31], the ring-radii optimization is performed before ring-spacing optimization. The ring-spacing optimization starts with the best result of the ring-radii optimization, effectively focusing primarily on homogeneity and not on the field strength.

The following iterative algorithm for the magnet design summarizes the discussion above:

#### Magnet design algorithm:

1. Initialization:
  - a. Specify target field vector  $\mathbf{b}_T$
  - b. Specify individual magnet strength  $M_0$ , number of layers  $L$  and the spacings  $\kappa$  and  $\delta$  between the magnets within a ring
  - c. Specify the number of rings  $N$
2. Initial guess:  $\mathbf{r}_0 = \mathbf{r}_{\min}$
3. Update ring-radii:  $\mathbf{r}_{i+1} = \mathbf{r}_i + \mathcal{J}_r^+ [\mathbf{b}_T - \mathbf{b}_z(\mathbf{r}_i)]$
4. If  $\frac{\|\mathbf{b}_T - \mathbf{b}_z(\mathbf{r}_i)\|}{\|\mathbf{b}_T\|} \leq \varepsilon_r$ : continue, else: repeat step 3
5. Mirror rings, check for improvement
6. Make initial guess  $\mathbf{d}_0 = \mathbf{d}_{\min}$
7. Update ring-spacing:  $\mathbf{d}_{i+1} = \mathbf{d}_i + \mathcal{J}_x^+ [\mathbf{b}_T - \mathbf{b}_z(\mathbf{d}_i)]$
8. If  $\frac{\|\mathbf{b}_T - \mathbf{b}_z(\mathbf{d}_i)\|}{\|\mathbf{b}_T\|} \leq \varepsilon_z$ : stop, else: repeat step 6
9. Mirror rings, and check for improvement

## Power optimised gradient coils

The gradient coils are designed using a previously implemented (quasi-static) target field approach [38–41]. The most important equations are summarized below, as well as the design choices which are integrated into the framework.

The target field method finds a continuous current density represented by a weighted sum of sinusoidal basis functions

which create a prescribed gradient field where the same FOV is used as for the magnet design. The wire patterns are found by taking contours of the corresponding stream function [44]. In particular, the stream function presented by Forbes and Crozier [40], namely,

$$\begin{aligned} \psi(\varphi, x) = & - \sum_{p=1}^P \frac{2\Lambda}{p\pi} \gamma_p \cos\left(\frac{p\pi(x+\Lambda)}{2\Lambda}\right) \\ & + \sum_{p=1}^P \sum_{q=1}^Q \frac{2\Lambda}{p\pi} [\alpha_{pq} \cos(q\varphi) + \beta_{pq} \sin(q\varphi)] \sin\left(\frac{p\pi(x+\Lambda)}{2\Lambda}\right) \end{aligned} \quad (7)$$

is used for gradient design. The above equation is defined for  $-\Lambda < x < \Lambda$ , where  $2\Lambda$  is the length of the coil, and  $\alpha_{pq}$ ,  $\beta_{pq}$  and  $\gamma_p$  are unknown expansion coefficients.  $P$  and  $Q$  determine the number of harmonic modes that are taken into account and if this number is sufficiently large then the sum of the trigonometric basis functions weighted by the expansion coefficients gives an accurate representation of the surface current for any chosen target field. The ideal weighting is found by substituting the corresponding equations for the current density into Biot-Savart's law and rewriting the resulting expressions as a system of equations for the unknown expansion coefficients. The desired expansion coefficients are obtained by solving this typically ill-conditioned system (see Eq. (8))

Using prior knowledge about the desired coil structure and the target fields, the number of unknowns (number of expansion coefficients) can be significantly reduced. Specifically, to realize  $x$ - and  $z$ -gradient coils, only the expansion coefficients  $\alpha_{pq}$  are required, while for  $y$ -gradients only the  $\beta_{pq}$  coefficients need to be taken into account.

As mentioned above, the system of equations for the expansion coefficients is typically ill-conditioned and requires regularisation. Specifically, for each gradient coil, the expansion coefficients are determined by minimizing the regularised least squares functional

$$\mathcal{F}(\mathbf{a}) = \frac{1}{2} \|\mathbf{W}\mathbf{a} - \mathbf{k}\|_2^2 + \frac{1}{2} \lambda \mathbf{a}^T \Upsilon \mathbf{a}. \quad (8)$$

Here,  $\mathbf{W}$  is the system matrix representing the Biot-Savart relationships of the transverse field component,  $\mathbf{a}$  is the vector containing the unknown expansion coefficients,  $\mathbf{k}$  is the target field vector and  $\Upsilon$  is a diagonal matrix containing coil efficiency parameters described in [41]. In case of gradient coil design, the regularisation parameter  $\lambda$  controls the trade-off between gradient efficiency ( $[\text{TA}^{-1} \text{m}^{-1}]$ ) and how well the desired linear target field is produced. Performing a sweep of this regularization parameter results in gradient coils with different values for these two performance metrics. The maximum linearity error is determined using.

$$\sigma = \max \left\{ \left| \frac{G_T(\mathbf{x}) - G(\mathbf{x})}{G_T(\mathbf{x})} \cdot 100 \right| \right\}. \quad (9)$$

Where,  $G_T(\mathbf{x})$  is the targeted field and  $G(\mathbf{x})$  is the realised field. In general, the efficiency and linearity error increase with higher values of  $\lambda$ , choosing a maximum allowable error thus results in the most efficient coil. The number of turns influence the gradient efficiency. This number is taken such that a minimum specified wire spacing is reached. The value of the gradient target field strength can be chosen arbitrarily as the resulting value of the stream function and thus the required current going through the wires will scale with it.

The maximum length and radius of the gradient coils are determined by the dimensions of the magnet. The  $y$ - and  $z$ -gradients have the same wire patterns but are rotated by  $45^\circ$  with respect to each and other. Due to the target field structure combined with the coil geometry these gradient coils are intrinsically more efficient than the  $x$ -gradient coils, and require a reduced length to obtain a specified linearity. Given these observations, the  $x$ -gradient is chosen to be the inner coil followed by the  $y$ - and  $z$ -gradients. It can occur that, depending on the number of magnet layers and the target field strength, one or both of the transverse gradient coils do not fit inside the magnet. In this case it is possible to place the gradient coil outside the magnet structure, which has been proposed in [45], but brings a significant efficiency penalty.

Finally, the gradient coils resulting from this method do not require active shielding due to the absence of a cryostat which is conventionally the main structure in which eddy currents are induced. Wire wound gradient coils are chosen, due to their relative ease of construction using 3D printed moulds. The following design algorithm is followed for the gradient coils.

#### Gradient design algorithm:

1. Initialization:
  - a. Obtain gradient coil length and radius from magnet dimensions
  - b. Specify target gradient field vector  $\mathbf{k}$
  - c. Specify the wire thickness, and minimum wire spacing
  - d. Specify maximum linearity
2. Find expansion coefficients  $\sim$  for a range of the regularization parameter  $\lambda$
3. Determine gradient efficiencies and linearities for all values of  $\lambda$
4. Find most efficiency coil corresponding to the maximum allowable linearity error.

## Target field RF coils

The same approach taken for the gradient coils is used to obtain wire patterns for the RF coils. A quasi-static field approach has proven to be appropriate for the design of the field pattern [45]. A solenoid with distributed wires is obtained by taking a uniform target field throughout the volume with the field directed along the axis of the bore. Only the first term  $\gamma_p$  from Eq. (7) needs to be taken into account. Subsequently, the functional Eq. (8) can be constructed, with  $W$  now containing the Biot-Savart relationships corresponding to the axial field direction. The regularisation parameter controls a trade-off between the transmit efficiency and how well the field is produced within the target region. The user specifies a certain uniformity error, which results in the particular wire pattern. By specifying the target volume and the allowable RF uniformity error, the resulting field, though a quasi-static simulation, is representative for a  $B_1^+$  pattern. The frequency-spread in the  $B_0$  field gives an indication of the coil bandwidth that can be obtained while still exciting all frequencies in the target volume. The finite support of the hard RF pulse results in a sinc function in the frequency domain. The FWHM of the central lobe of the sinc function should cover the spread in frequencies created by the  $B_0$  field inhomogeneities. The minimum desired pulse length defined by the user is used to check if the flip angles can be achieved given the simulated transmit efficiency and the RF amplifier capabilities. Hard RF pulses used for volume excitations and refocussing are considered and the maximum desired flip angle is set at 180 degrees. The transmit efficiency is computed using 1 mm copper wire.

## Image spatial resolution

The effective spatial resolution of the image is represented by the point spread function (PSF) which is different in the frequency and phase encoding directions. Here we assume that three-dimensional images are acquired using one frequency and two phase encoding gradients. The PSF in the frequency encoding direction depends upon the maximum gradient strength, the  $B_0$ -field inhomogeneity, and the digital resolution (FOV divided by the number of data points acquired): in the phase encoding directions the digital resolution is the only factor that needs to be considered. The resolution in the frequency encoding direction is affected by the exponential  $T_2^*$  decay, which has a blurring effect. The linewidth is assumed to be a Lorentzian, such that  $\frac{1}{T_2^*} = \frac{1}{T_2} + \gamma \Delta B_0$  [46], where  $T_2$  is the transverse relaxation time, and the second term consists of the product of the gyromagnetic ratio  $\gamma$  and  $\Delta B_0$ , the voxel field variation. The assumption is made that the field inhomogeneities dominate ( $\gamma \Delta B_0 \gg \frac{1}{T_2}$ ), such that  $T_2^* \approx \frac{1}{\gamma \Delta B_0}$ . This assumption is valid

for typical  $T_2$  times encountered during *in-vivo* experiments combined with the inhomogeneities of typical Halbach systems which are hundreds to thousands of parts per million (ppm) [47]. Taking the sum of the individual phase differences caused by the inhomogeneities throughout the region of interest results in an exponential decay which goes to zero in the limit

$$\lim_{t \rightarrow \infty} \sum_x e^{-i\gamma[B_0 + \Delta B(x)]t} \rightarrow 0. \quad (10)$$

Its Fourier transform is a Lorentzian. Given the situation that the magnet field inhomogeneities are the dominant term,  $T_2^* \approx \frac{1}{\pi FWHM}$ , where FWHM is the full width half maximum of the Lorentzian. The effective resolution due to the  $T_2^*$  filtering effect [48] can be expressed as

$$\Delta z_{eff} = N_z \Delta z \frac{\tanh\left(\frac{\pi FWHM}{2BW}\right)}{1 - e^{-\frac{\pi FWHM N_z}{2BW}}} \quad (11)$$

where  $BW$  is the maximum obtainable readout bandwidth of the gradient system determined with the least efficient gradient,  $\Delta z$  the digital resolution and  $N_z$  the number of points.

### Example systems

To demonstrate the potential of the framework, two distinct systems are designed. The first is a system suitable for adult neuroimaging with input parameters guided by values from the literature [10, 12, 18, 28, 32]: FOV =  $200 \times 200 \times 200$  mm<sup>3</sup>, isotropic effective spatial resolution  $\sim 3 \times 3 \times 3$  mm<sup>3</sup>, data matrix  $80 \times 66 \times 66$  corresponding to  $0.5 \pm 0.1$  mm PSF blurring in the readout direction. The minimum field strength is set to 45 mT, slightly lower than previous systems. To ensure the head can reach the centre of the magnet, taking into account the shoulders of the subject, the maximum bore length cannot exceed 550 mm and the minimum bore diameter is set at 300 mm. This ensures sufficient space for the RF coil which has a diameter and length of 240 mm, equivalent to solenoidal coils designed for previous low-field systems [9, 10, 32].

The second system is designed for imaging of the hand and wrist. A cylindrical target volume of length 200 mm and diameter 120 mm is used. The RF coil has a length of 240 mm and a diameter of 140, the same ratio with respect to the target volume as the neuroimaging system. Based on previous literature [9, 33] we used a target isotropic resolution of 1.75 mm, obtained with a  $200 \times 120 \times 120$  mm<sup>3</sup> FOV and  $125 \times 68 \times 68$  datapoints, i.e. a maximum  $0.15 \pm 0.01$  mm blurring PSF in the readout direction is allowed. The minimum field strength is set at 80 mT. The maximum length of the system is set to 500 mm to allow positioning of the wrist at the centre of the magnet.

Both systems are designed using identical RF and gradient amplifiers, again with specifications based on previous designs [41]. The RF amplifier has a gain of 54 dB combined with a spectrometer which can output a maximum 1 mW. The gradient amplifier has a maximum current output of 30 A. The maximum gradient linearity error is set at 5% throughout the target region. The maximum allowable RF non-uniformity is set at 20% and the minimum pulse length to 50  $\mu$ s. The individual cubical magnets used have a remanence of 1.3 T and are 12 mm<sup>3</sup>. A summary of the input parameters used for the design of both systems are shown in Table 1.

## Results

### Neuroimaging system

The initial magnet setup consists of 25 rings and 20 mm ring-spacing, resulting in a length/diameter ratio of 1.7, which is an appropriate initial estimate based on previous designs [28, 31]. In the first framework iteration a homogeneity of 4329 ppm after the ring-distance optimisation is reached. Combining this with the gradient efficiency of the least efficient  $x$ -gradient ( $0.24 \text{ mTm}^{-1}\text{A}^{-1}$ ), and a 30 A maximum current output, results in an effective spatial resolution of 10.3 mm. This is well above the specified 3 mm and is due to poor magnet homogeneity (4329 ppm) caused by the small length/diameter ratio. An additional ring is added to improve the homogeneity. The resulting 26 rings produce a more homogenous design (1075 ppm). The longer magnet allows the length of the weakest ( $x$ -) gradient to also increase, which improves its efficiency slightly from 0.24 to  $0.26 \text{ mTm}^{-1}\text{A}^{-1}$ . This results in an effective resolution of 4.1 mm, which is still above the pre-defined maximum. An additional ring, 27 rings in total, increases the magnet homogeneity to 381 ppm and the increase in length allows for a longer gradient coil which again increases its efficiency such that the effective resolution equals 2.6 mm, which is within the specifications. The next round of iterations determines if the static magnetic field strength can be increased to improve the SNR while still maintaining the required spatial resolution: in each iteration the field strength is increased in steps of 1 mT. Consequently, the magnet length and average diameter are decreased, making the system more compact. For this example, iteration seven of the framework gives the most compact configuration with the strongest magnetic field (49 mT) and sufficient homogeneity to obtain an effective resolution of 3 mm. Increasing the magnetic field even further would cause the effective resolution to exceed the specified maximum.

The 240 mm length and diameter solenoid has a uniformity within the FOV of 20% and a transmit efficiency of 26



$\mu\text{T/W}$ . Given the 1 mW output of the spectrometer and the RF amplifier gain of 54 dB, a maximum desired flip angle of 180 degrees and a pulse duration of 50  $\mu\text{s}$ , the  $B_1$ -field required equals 234  $\mu\text{T}$  and therefore 9 W is required, which is well below the maximum output power. The final check is that the 50  $\mu\text{s}$  RF pulse excites a bandwidth much greater than the  $B_0$  inhomogeneity, which indeed is the case.


Figure 3 shows the final output of the framework, where the gradient coils, the magnet, and the RF coil are shown separately. Table 2 shows the output parameters of the framework for every iteration.

### Extremity system

As the bore size is 1.5 times smaller than the previous example, but the volume of interest is now a cylinder which has a length relatively long compared to the diameter, the initial guess for the number of rings is chosen to be 23 with each ring holding two layers of magnets. During the first framework iteration it becomes clear that the specified field strength of 80 mT requires the centre ring-diameters to become so small that the outer gradient coil ( $z$ -gradient coil) does not fit inside the magnet. Due to the intrinsic efficiency difference between the  $x$ - and  $y/z$ -gradients it is feasible to place the  $z$ -gradient on the exterior of the magnet.

The efficiency of the  $z$ -gradient is  $1.5 \text{ mTm}^{-1}\text{A}^{-1}$ , while the  $x$ -gradient coil has an efficiency of  $0.74 \text{ mTm}^{-1}\text{A}^{-1}$  and is thus still the least efficient. The homogeneity of the magnet is 1208 ppm which results in an effective resolution of 1.88 mm, obtained with 23 rings. Since the target is 1.75 mm, the homogeneity and/or the gradient efficiency needs to be improved. This is realized by including an additional ring, which increases the homogeneity to 216 ppm and the gradient efficiency of the least efficient gradient from 0.74 to  $0.82 \text{ mTm}^{-1}\text{A}^{-1}$ . The effective resolution with 24 rings is 1.62 mm, which is better than required. In the successive framework iterations, the target field is increased until the value of 84 mT is reached which gives an effective resolution of 1.75 mm, computed with a homogeneity of 803 while the gradient efficiency goes from 0.82 to  $0.83 \text{ mTm}^{-1}\text{A}^{-1}$ .

The spectrometer characteristics are the same as for the previous case. The transmit efficiency of the RF coil, designed with a diameter of 140 mm is  $58 \mu\text{T/W}$ , thus 6.5 W is required for a  $180^\circ$  pulse. A 50  $\mu\text{s}$  pulse excites the entire imaging FOV. Relevant output parameters of the framework for every iteration are shown in Table 3. Figure 4 shows the gradient magnet and RF coils at the final stage of the framework.



**Fig. 3** Neuro system:  $x$ -gradient: black,  $y$ -gradient: yellow,  $z$ -gradient: red, RF coil (green) and magnet corresponding to the final iteration of the system design framework. A single layer of magnets is shown for clarity. The values on the axes indicate the maximum dimensions

**Table 2** Framework output parameters per iteration neuroimaging system

Framework iteration	1	2	3	4	5	6	7
<b>Magnet parameters</b>							
N rings	25	26	27	27	27	27	27
$B_0$ [mT]	45	45	45	46	47	48	49
Magnet length [mm]	493	512	529	529	532	535	528
Ring radii optimisation [ppm]	6093	4773	3303	2861	3116	3210	4023
Ring spacing optimisation [ppm]	4329	1075	381	120	420	685	878
Number of magnets	3368	3532	3705	3617	3554	3485	3306
<b>Gradient coil parameters</b>							
Length x [mm]	493	512	529	529	529	529	528
Length y [mm]	437	457	475	475	475	475	450
Length z [mm]	457	457	457	457	457	457	450
Radius x [mm]	150	150	150	150	150	150	150
Radius y [mm]	155	155	155	155	155	155	155
Radius z [mm]	160	160	160	160	160	160	160
Efficiency $\eta_x$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	0.24	0.26	0.31	0.31	0.31	0.31	0.31
Efficiency $\eta_y$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	1.1	1.2	1.3	1.3	1.3	1.3	1.2
Efficiency $\eta_z$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	0.93	0.93	0.93	0.93	0.85	0.85	0.98
<b>Imaging parameters</b>							
FWHM [Hz]	1989	568	61	51	43	88	263
Readout resolution [mm]	10.4	4.1	2.6	2.6	2.6	2.7	3.0
<b>RF coil parameters</b>							
Transmit Efficiency [ $\mu\text{T}/\text{W}$ ]							26
Required $B_1^+$ : $\alpha = 180^\circ$ [ $\mu\text{T}$ ]							234
Maximum pulse length [ $\mu\text{s}$ ]							545

## System comparison

The total computation time to reach the above discussed configurations is less than 15 min. A single Newton iteration takes  $\sim 5$  s given the parameters used in the examples and using standard computing hardware (Windows, intel i3, 4-cores with 8 GB of RAM).

Figure 5 shows a side-view of the final configuration for both the neuroimaging and extremity system. The top figure corresponding to the neuro system shows that the gradient coils fit inside the bore, while in the bottom figure the outer coil ( $z$ -gradient coil) needs to be placed at the exterior of the bore. This is because the homogeneity of the head system is restricted by the maximum length of the system, adding additional rings would create a more homogenous system, but will make it impossible for the head to reach the centre of the magnet. The extremity system requires one of the gradient coils to be placed exterior to the magnet in order to design a small enough magnet to reach 80 mT. An alternative approach would be to add an additional layer of magnets. However, this would increase the weight of the magnet significantly.

Figure 6 shows the linearity vs gradient efficiency plots for the final designs of the two systems. The curve of the


$x$ -gradient coil is the steepest meaning that an increase in efficiency is the most costly in terms of linearity. The gradient coils of the wrist system obtain the specified linearity more easily due to the dimensional difference between the systems.

## Discussion


The described framework gives the designer insight into how the magnet design influences the optimal gradient and RF geometries and consequently the imaging parameters that can be obtained with a specific setup. This is achieved with the integration of the time-efficient computational methods which determine the magnet, gradient and RF geometries. Integration has been shown to be important for dealing with the varying magnet ring-radii, which constrain the design space for the gradient coils and influences the obtainable spatial resolution. The two examples in the previous section show that systems for typical low-field point-of-care applications can be successfully designed, with the design of the neuro-imaging system similar in size and field strength to an earlier system [31]. The extremity system shows the importance of integrating

**Table 3** Framework output parameters per iteration extremity syste

Framework iteration	1	2	3	4	5	6
<b>Magnet parameters</b>						
N rings	23	24	24	24	24	24
$B_0$ [mT]	80	80	81	82	83	84
Magnet length [mm]	448	469	469	468	467	436
Ring-radii optimisation [ppm]	5683	2758	3035	2919	4399	3825
Ring-spacing optimisation [ppm]	1208	216	352	563	363	803
Number of magnets	1904	1990	1972	1960	1938	1922
<b>Gradient coil parameters</b>						
Length x [mm]	448	469	448	448	448	448
Length y [mm]	374	370	370	370	320	327
Length z [mm]	400	400	400	400	400	400
Radius x [mm]	105	105	105	105	105	105
Radius y [mm]	110	110	110	130	110	110
Radius z [mm]	154	154	152	151	149	147
Efficiency $\eta_x$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	0.74	0.82	0.82	0.82	0.83	0.83
Efficiency $\eta_y$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	2.5	2.6	2.6	2.6	2.0	2.1
Efficiency $\eta_z$ [ $\text{mTm}^{-1}\text{A}^{-1}$ ]	1.5	1.5	1.5	1.5	1.5	1.6
<b>Imaging parameters</b>						
FWHM [Hz]	332	37	32	84	127	214
Readout resolution [mm]	1.88	1.62	1.61	1.66	1.68	1.75
<b>RF coil parameters</b>						
Transmit Efficiency [ $\mu\text{T/W}$ ]						58
Required $B_1^+$ : $\alpha = 180^\circ$ [ $\mu\text{T}$ ]						234
Maximum pulse length [ $\mu\text{s}$ ]						348

**Fig. 4** Extremity system: x-gradient: black, y-gradient: yellow, z-gradient: red), RF coil (green) and magnet corresponding to the final iteration of the system design framework. A single layer of magnets is

shown for clarity. The z-gradient coil (red) is located at the exterior of the magnet. The values on the axes indicate the maximum dimensions




**Fig. 6** Linearity vs gradient efficiency plots for the three gradient coils. Linearity computed with respect to linear field inside the volume of interest. Red  $z$ -gradient, yellow  $y$ -gradient and black  $x$ -gradient

the design methods, as the placement of one gradient outside the magnet allows sufficiently small ring-diameters to obtain the targeted field strength with a two-layer magnet design, while the combined performance of the gradient system and magnet homogeneity are sufficient to obtain the desired spatial resolution.

One of the strengths of the framework is that the target field methods described are compatible with any target region. The shape, location and values are the designer's choice, the magnet, gradient and RF coils are all optimised for the specified region. This includes asymmetric designs which are not considered in this work but are within the framework's capabilities.

In terms of the magnet design, the combination of ring-radii and -separation optimisation using a Newton algorithm shows the potential to be effective in obtaining highly homogenous magnet designs within short computational times compared to other numerical methods such as genetic algorithms. Some of the inputs requested from the designer, such as the starting number of rings, require some prior knowledge. This choice affects the amount of framework iterations necessary, but the initial conditions do not influence the final result, unless the number of starting rings is chosen larger than the minimum length solution, in which case the framework would finish after a single iteration. Another design choice is the  $B_0$  field strength, a sense for this parameter with a specified configuration can be obtained by solving the forward problem (3) using the minimum allowable radius. This will result in the maximum obtainable field strength of this configuration and gives the designer insight into the number of magnet layers required and the individual magnet strength to use.



**Fig. 5** Side view of two magnet systems showing how the gradient coils fit into the design. The top figure corresponds to the final iteration of the neuroimaging system. The bottom corresponds to the extremity system. The blocks represent the magnet positions, red  $z$ -gradient, yellow  $y$ -gradient, black  $x$ -gradient. The top figure has a conventional layout, where the gradients are on the inside of the magnet. In the bottom figure the  $z$ -gradient is placed outside the magnet due to insufficient space on the inside

In this work two magnet designs with a homogenous  $B_0$  field are discussed, it is also possible to consider a linear  $B_0$  field, effectively replacing one of the gradient coils. For a Halbach system the  $y$ - or  $z$ -gradient can be replaced by changing the fixed orientation of the magnets [27, 28].

The side views of the magnet rings in Fig. 5 show that the outer rings have reached their minimum spacing. If this constraint is removed, the model would produce overlapping rings. This indicates that an increase in magnetisation towards the ends of the magnet can further improve the homogeneity. One can consider adding an additional layer of magnets to the outer rings when this occurs [49].



Manufacturing errors have known to influence the realised homogeneity of Halbach systems [23, 28, 31]. There are a number of shimming methods that can be used to obtain the desired homogeneity [50, 51]. Consequently, the framework gives the homogeneity that needs to be obtained after shimming. Previous work has shown us that the gradient coil construction errors have less of an influence and that the magnetic fields modelled correspond closely with the measured fields created by the constructed coils [38].

In terms of gradient coil design there are alternative approaches which can optimize the gradient coil inductance [36]: requirements regarding switching speeds can easily be added into the framework. However, the greater challenge for compact, point-of-care systems with relatively low gradient amplifier current is typically to obtain sufficient homogeneity and linearity. Therefore, the emphasis is on optimising the power efficiency of the coils for a given linearity.

The blurring in the readout direction is determined with the least efficient gradient coil, to simulate the worst-case scenario and give the user flexibility in the choice of the readout-direction. In most cases the  $x$ -gradient is least efficient and therefore the choice could be made to design a system where the readout direction is always taken transverse to the bore. Consequently, the homogeneity and/or gradient amplifier requirements can be relaxed significantly.

The framework can be extended to include the effects of concomitant fields. These undesired vector components are unavoidable when creating the desired gradient fields, and can create distortions when the gradient field strength becomes significant with respect to  $B_0$  field strength [52–55]. One way to consider these effects is to constrain the dimensionless figure of merit  $\epsilon = \frac{G_{max}r_{max}}{B_0}$ , which gives an indication of the amount of distortions to expect. A more advanced way would be to consider the blurring and warping effects separately and take these into account when calculating the effective resolution and warping due to gradient non-linearities. Most models only account for the concomitant components of linear fields. Therefore, linearity must be constrained to a lower percentage for the models to be accurate, which can change the geometry of the coils. These effects are pulse sequence dependent and no research has been yet been published on concomitant fields specific for Halbach systems with cylindrical gradient coils.

The gradient coil heating is not modelled, but is an important factor that needs to be considered by the designer when choosing the wire thickness since excess heating can influence the strength of the permanent magnets or even damage them. There are a number of example systems in literature that can be referenced which use wire wound gradients [9, 10, 18, 28, 32, 33, 38, 41, 45].

The main RF coil considered in this work is the solenoidal coil, these coils have proven themselves for transverse

magnetic field systems due to their superior sensitivity with respect to the saddle coil, which is the other single channel coil to consider [35]. The saddle coil however, has a better uniformity for a fixed length, which can be beneficial for short RF-coils.

**Data availability** All data and code that support the findings of this study are available from the author, upon request.

**Conflict of interest** The authors declare no conflicts of interest.

**Ethical standard** This article does not contain any studies with human participants performed by an of the authors.

**Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

## References

1. Anazodo UC et al (2022) A framework for advancing Sustainable Magnetic Resonance Imaging Access in Africa. *NMR Biomed* 11:4846
2. Ogbole GI, Adeyomoye AO, Badu-Peprah A, Mensah Y, Nzeh DA (2018) Survey of magnetic resonance imaging availability in West Africa. *Pan Afr Med J* 30:240
3. International atomic Energy Agency, MRI units per million inhabitants. [Online]. Available: <https://humanhealth.iaea.org/HHW/DBStatistics/IMAGINEMaps3.html>
4. Hori M, Hagiwara A, Goto M, Wada A, Aoki S (2021) Low-field magnetic resonance imaging: Its History and Renaissance. *Investig Radiol* 56(11):669–679
5. Marques JP, Simonis FFJ, Webb AG (2019) Low-field MRI: An MR Physics Perspective. *J Magn Reson Imaging* 49:1528–1542. <https://doi.org/10.1002/jmri.26637>
6. Sarracanie M, Salameh N (2020) Low-field MRI How low can we go? A fresh view on an old debate. *Front Phys*. <https://doi.org/10.3389/fphy.2020.00172>
7. Huang S, Ren ZH, Obruchkov S, Gong JI, Dykstra R, Yu W (2019) Portable low-cost MRI system based on permanent magnets/magnet arrays. *Investig Magn Reson Imaging* 23:179
8. Wald LL, McDaniel PC, Witzel T, Stockmann JP, Cooley CZ (2019) Low-cost and portable MRI. *J Magn Reson Imaging* 52:686–696
9. Nakagomi M, Kajiwara M, Matsuzaki J, Tanabe K, Hoshiai S, Okamoto Y, Terada Y (2019) Development of a small car-mounted magnetic resonance imaging system for human elbows using a 0.2 T permanent magnet. *J Magn Reson* 304:1–6
10. He Y, He W, Tan L, Chen F, Meng F, Feng H, Xu Z (2020) Use of 2.1 MHz MRI scanner for brain imaging and its preliminary results in stroke. *J Magn Reson* 319:106829
11. Sheth KN, Mazurek MH, Yuen MM et al (2021) Assessment of brain injury using portable, low-field magnetic resonance imaging at the bedside of critically ill patients. *JAMA Neurol* 78:41



12. Mazurek MH et al (2021) Portable, bedside, low-field magnetic resonance imaging for evaluation of intracerebral hemorrhage. *Nat Commun.* <https://doi.org/10.1038/s41467-021-25441>
13. Cahn B et al (2020) Deployment of Portable, Bedside, Low-field Magnetic Resonance Imaging for Evaluation of Stroke Patients. *Neurology* 94(15):272
14. Deoni SCL et al (2021) Accessible pediatric neuroimaging using a low field strength MRI scanner. *Neuroimage* 238:118273
15. Deoni SC et al (2022) Development of a mobile low-field MRI scanner. *Sci Rep.* <https://doi.org/10.1038/s41598-022-09760-2>
16. Chetcuti K et al (2022) Implementation of a low-field portable MRI scanner in a resource-constrained environment: Our experience in Malawi. *AJNR Am J Neuroradiol.* <https://doi.org/10.3174/ajnr.a7494>
17. Arnold TC, Baldassano SN, Litt B, Stein JM (2021) Simulated diagnostic performance of ultra-low-field MRI: Harnessing open-access datasets to evaluate novel devices. *Magn Reson Imaging.* <https://doi.org/10.1101/2021.07.02.21259789>
18. Liu Y et al (2021) A low-cost and shielding-free ultra-low-field brain MRI scanner. *Nat Commun.* <https://doi.org/10.1038/s41467-021-27317-1>
19. Cooley CZ et al (2020) Design and implementation of a low-cost, tabletop MRI scanner for education and Research Prototyping. *J Magn Reson* 310:106625
20. Halbach K (1979) Strong rare earth cobalt quadrupoles. *IEEE Trans Nucl Sci* 26:3882–2884
21. Halbach K (1980) Design of permanent multipole magnets with oriented rare earth cobalt material. *Nucl Instrum Methods* 169:1–10
22. Halbach K (1982) Perturbation effects in segmented rare earth cobalt multipole magnets. *Nucl Instrum Methods* 198:213–215
23. Raich H, Blümmler P (2004) Design and construction of a dipolar halbach array with a homogeneous field from identical bar magnets: NMR Mandhalas Concepts Magn. Reson, Part B 23B:16–25
24. Windt CW, Soltner H, Dvan D, Blümmler P (2011) A portable halbach magnet that can be opened and closed without force: The NMR-cuff. *J Magn Reson* 208:27–33
25. Soltner H, Blümmler P (2010) Dipolar Halbach magnet stacks made from identically shaped permanent magnets for Magnetic Resonance. *Concepts Magn Reason Part A* 36A:211–222
26. Blümmler P (2016) Proposal for a permanent magnet system with a constant gradient mechanically adjustable in direction and strength. *Concepts Magn Reason Part B* 46:41–48
27. Cooley CZ, Haskell MW, Cauley SF, Sappo C, Lapierre CD, Ha CG, Stockmann JP, Wald LL (2018) Design of sparse halbach magnet arrays for portable MRI using a genetic algorithm. *IEEE Trans Magn* 54:1–12
28. Cooley CZ et al (2020) A portable scanner for Magnetic Resonance Imaging of the brain. *Nat Biomed Eng* 5:229–239
29. Srinivas SA et al (2021) External dynamic interference estimation and removal (EDITER) for low Field MRI. *Magn Reson Med* 87:614–628
30. Mateen FJ, Cooley CZ, Stockmann JP, Rice DR, Vogel AC, Wald LL (2021) Low-field portable brain MRI in CNS demyelinating disease. *Mult Scler Relat Disord* 51:102903
31. O'Reilly T, Teeuwisse WM, Webb AG (2019) Three-dimensional MRI in a homogenous 27 cm diameter bore Halbach Array Magnet. *J Magn Reson* 307:106578
32. O'Reilly T, Teeuwisse WM, Gans D, Koolstra K, Webb AG (2020) In vivo 3D brain and extremity MRI at 50 MT using a permanent magnet halbach array. *Magn Reason Med* 85:495–505
33. Guallart-Naval T et al (2022) Benchmarking the performance of a low-cost magnetic resonance control system at multiple sites in the open marcos community. *NMR Biomed.* <https://doi.org/10.1002/nbm.4825>
34. Hidalgo-Tobon SS (2010) Theory of gradient coil design methods for magnetic resonance imaging. *concepts magn. Reson A: Bridg Educ Res* 36A:223–242. <https://doi.org/10.1002/cmr.a.20163>
35. Hoult DI, Richards RE (1976) The signal-to-noise ratio of Nuclear Magnetic resonance experiment. *J Magn Reson* 24(1):71–85
36. Turner R (1988) Minimum inductance coils. *J Phys E Sci Instrum* 21:948–952
37. Turner R (1986) A target field approach to optimal coil design. *J Phys D: Appl Phys.* <https://doi.org/10.1088/0022-3727/19/8/001>
38. de Vos B, Fuchs P, O'Reilly T, Webb A, Remis R (2020) Gradient coil design and realization for a Halbach-based MRI system. *IEEE Trans Magn* 56:1–8
39. Forbes LK, Crozier S (2001) A novel target-field method for finite-length magnetic resonance Shim Coils: I. Zonal Shims. *J Phys D: Appl Phys* 34:3447–3455
40. Forbes LK, Crozier S (2002) A novel target-field method for finite-length magnetic resonance Shim Coils: II. Tesseral Shims. *J Phys D: Appl Phys* 35:839–849
41. de Vos B et al (2021) Design, characterisation and performance of an improved portable and sustainable low-field MRI system. *Front Phys.* <https://doi.org/10.3389/fphy.2021.701157>
42. Petruska AJ, Abbott JJ (2013) Optimal Permanent-magnet geometries for dipole field approximation. *IEEE Trans Magn* 49:811–819
43. Nocedal J, Wright SJ (2006) Line search methods. In: Priya D (ed) numerical optimisation, 2nd edn. Springer, New York, pp 30–63
44. Brideson MA, Forbes LK, Crozier S (2001) Determining complicated winding patterns for shim coils using stream functions and the target-field method. *Concepts Magn Reson* 14:9–18
45. McDaniel PC (2020) Computational design and fabrication of portable mri systems dissertation. department of electrical engineering and computer science massachusetts institute of technology, Cambridge Massachusetts
46. Liang Z, Signal LPC (1999) Characteristics. In: Priya D (ed) Principles magnetic resonance imaging: a signal processing perspective. John Wiley & sons, pp 111–114
47. O'Reilly T, Webb AG (2021) In vivo T1 and T2 relaxation time maps of brain tissue, skeletal muscle, and lipid measured in healthy volunteers at 50 mT. *Magn Reason* 87:884–895
48. Brown RW, Cheng YN, Haacke EM, Thompson MR, Venkatesan R (2014) Filtering and Resolution in Fourier Transform Image Reconstruction. In: Priya D (ed) Magnetic Resonance Imaging, physical principles and sequence design, 2nd edn. Wiley Blackwell, New Jersey, p 281
49. Purchase AR et al (2021) A short and light, sparse dipolar halbach magnet for MRI. *IEEE Access.* 9:95294
50. Parker AJ, Zia W, Rehorn CWG, Blümich B (2016) Shimming halbach magnets utilizing genetic algorithms to profit from material imperfections. *J of Magn Reason* 265:83–89
51. Wenzel K, Alhamwey H, O'Reilly T, Riemann LT, Silemek B, Winter L (2021) B0-shimming methodology for affordable and compact low-field magnetic resonance imaging magnets. *Front Phys.* <https://doi.org/10.3389/fphy.2021.704566>
52. Volegov PL, Mosher JC, Espy MA, Kraus RH Jr (2005) On concomitant gradients in low-field MRI. *J Magn Reson* 175(1):103–13
53. Zhou XJ, Tan SG, Bernstein MA (1998) Artifacts induced by concomitant magnetic field in fast spin-echo imaging. *Magn Reson Med* 40(4):582–591
54. Bernstein MA, Zhou XJ, Polzin JA, King KF, Ganin A, Pelc NJ, Glover GH (1998) Concomitant gradient terms in phase contrast MR: analysis and correction. *Magn Reson Med* 39(2):300–308
55. King KF, Ganin A, Zhou XJ, Bernstein MA (1999) Concomitant gradient field effects in spiral scans. *Magn Reson Med* 41(1):103–12

**Publisher's Note** Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.